

Bond University
Research Repository



Does a Novice Technician Produce Results Similar to That of an Experienced DXA Technician When Assessing Body Composition and Bone Mineral Density?

Persson, Carl; Shiel, Flinn; Climstein, Mike; Furness, James

Published in:
International Journal of Sport Nutrition and Exercise Metabolism

DOI:
[10.1123/ijsnem.2018-0299](https://doi.org/10.1123/ijsnem.2018-0299)

Licence:
Free to read

[Link to output in Bond University research repository.](#)

Recommended citation(APA):
Persson, C., Shiel, F., Climstein, M., & Furness, J. (2019). Does a Novice Technician Produce Results Similar to That of an Experienced DXA Technician When Assessing Body Composition and Bone Mineral Density? *International Journal of Sport Nutrition and Exercise Metabolism*, 29(4), 1-19. Advance online publication. <https://doi.org/10.1123/ijsnem.2018-0299>

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

For more information, or if you believe that this document breaches copyright, please contact the Bond University research repository coordinator.

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20
21
22

Does a novice technician produce results similar to that of an experienced DXA technician when assessing body composition and bone mineral density?

Carl Persson¹, Flinn Shiel¹, Mike Climstein^{2,3,4}, James Furness^{1,2}.

1 Faculty of Health Sciences and Medicine, Bond University, Gold Coast, QLD 4226, Australia

2 Water Based Research Unit, Bond Institute of Health & Sport, Bond University, Gold Coast, QLD 4226

3 School of Health and Human Sciences, Southern Cross University, Lismore, NSW, Australia,

4 Exercise, Health & Performance Faculty Research Group, Faculty of Health Sciences, The University of Sydney, Sydney, NSW, Australia

Address for correspondence: Dr James Furness, PhD, Faculty of Health Sciences and Medicine, Bond University, Robina, QLD, Australia. Email: jfurness@bond.edu.au

Running title – Are novice DXA technicians results similar to experienced?

23 ABSTRACT

24 Dual energy X-ray absorptiometry is a commonly used clinical assessment tool for body
25 composition and bone mineral density, which is gaining popularity in athletic cohorts. Results
26 from body composition scans are useful for athletic populations to track training and nutritional
27 interventions, whilst bone mineral density scans are valuable for athletes at risk of developing
28 stress fractures due to low bone mineral density. However, no research has ascertained if a
29 novice technician (accredited but not experienced) could produce similar results to an
30 experienced technician. Two groups of recreational athletes were scanned, one by an
31 experienced technician, one by a novice technician. All participants were scanned twice with
32 repositioning between scans. The experienced technician's reliability (ICC 0.989 – 0.998,
33 percentage change in mean -0.01 – 0.10), precision (typical error as CV% 0.01 to 0.47. standard
34 error of measurement percentage 0.61% - 1.39%) and sensitivity to change (smallest real
35 difference percentage 1.70% - 3.85%) were similar, however superior, to those of the novice
36 technician. The novice technician results were: reliability (ICC 0.985 – 0.997, percentage
37 change in mean -0.03 – 0.23), precision (typical error as CV% 0.03 – 0.75%, standard error of
38 measurement percentage 1.06% - 2.12%) and sensitivity to change (smallest real difference
39 percentage 2.73% - 5.86%). Extensive experience whilst valuable is not a necessary
40 requirement to produce quality results when undertaking whole body dual energy X-ray
41 absorptiometry scanning.

42

43 KEYWORDS

44 Reliability; Precision; sensitivity to change;

45

46

47

48 INTRODUCTION

49 Low bone mineral density (BMD) and associated conditions such as osteoporosis and
50 osteopenia are health problems that annually costs over 830 million dollars in Australia and
51 osteoporosis is a significant cause of morbidity and mortality (Johnell et al., 2006; Watts et al.,
52 2013). The need to accurately and effectively measure whole body and segmental BMD led to
53 the development of the DXA scanner, which is now considered the gold standard for BMD and
54 body composition (Blake et al., 2007; Lewiecki, 2005). Low BMD (osteoporosis and
55 osteopenia) is a concern for the general population as well as athletic population, as low BMD
56 increases the risk of stress or fragility fractures while an athlete is actively training, competing
57 and later in life (Kelsey et al., 2007; Scofield et al., 2012). Reduced cortical mass can
58 predispose athletes to lower limb stress fractures, with the incidence rate being as high at 20%
59 annually in track and field athletes (Bennell et al., 1996). Additionally, it is recognised that
60 endurance athletes (female runners and swimmers, male cyclists) and athletes who did not
61 partake in loaded and/or impact activities and sports as teens are at a higher risk of having low
62 BMD and subsequently developing bony stress related conditions (Fredericson et al., 2005;
63 Tenforde et al., 2015). This is due to factors including female athlete triad and excessive time
64 spent in sport with low cortical stress leading to weakened bone strength (Chen et al., 2013;
65 Fredericson et al., 2005; Tenforde et al., 2015). Therefore, screening these athletes via DXA
66 can act as an injury prevention tool for early intervention. If stress fractures are not correctly
67 treated and healed, they can result in a reduction in performance, an increase in pain, a loss of
68 training time and medical expenses; subsequent development of a complete fracture, non-
69 union, chronic pain, increased recovery time and possibly disability (Chen et al., 2013;
70 Schnackenburg et al., 2011).

71

72 Additionally, DXA's ability to assess whole body and segmental body composition (BC)
73 including lean mass (LM), fat mass (FM) and bone mineral content (BMC) has become an
74 important tool in the measurement of BC and is used in clinical, sporting and research settings
75 and is considered the reference standard (Buckinx et al., 2018). In the sporting setting, it is
76 known that LM and FM impact physical performance and the risk of injury and illness (Duthie,
77 2006; Georgeson et al., 2011; Hagmar et al., 2013; Stewart, 2001). Therefore, it is common
78 practice among the professional sporting population to have regular BC assessments to track
79 the effectiveness of training or nutritional interventions as any small change to BC can impact
80 performance (Duthie, 2006).

81

82 The International Society for Clinical Densitometry (ISCD) recommends precise measures
83 during preparation and positioning of the participant of a DXA BC scan (ISCD, 2015). For
84 whole body analysis of BC it has been shown that sources of biological error in DXA results
85 include hydration, stomach content and food consumption, time of day of scanning and pre-
86 scan physical activity (Hangartner et al., 2013; Nana et al., 2012, 2013). Furthermore sources
87 of technical error include artifacts such as clothing, number of technicians used to complete
88 scans and position of participant (Hangartner et al., 2013; Kiebzak et al., 2000; Kerr et al.,
89 2016; Nana et al., 2012). Given the importance of positioning, it is crucial the DXA technician
90 adheres to established best practice to ensure the most accurate and reliable results.

91

92 It has been reported that up to 64% of scans were deemed inadequate as they did not provide
93 sufficient accuracy when automatic analysis was applied and that manual analysis should be
94 undertaken, therefore the skill of the DXA scanning technician is vitally important (Baniak et
95 al., 2014). However, to date no research has focused on whether the experience of the DXA
96 scanner influences BC and BMD results. Kim et al. (2014) suggest a DXA technologist is

97 sufficiently experienced after performing repeated training in which the technologist
98 undertakes adjusting patient positioning, device manipulation and result analysis on 100
99 patients.

100

101 The increasing popularity to use DXA to assess and track change in BC and BMD over time
102 has created a larger need for qualified DXA technicians. As such, Australian (Kerr et al., 2016)
103 and USA universities (Standorth et al., 2016; Trexler et al., 2018) possess DXA scanners to
104 conduct research assessing BC and BMD in athletic/non-athletic and clinical (Newton et al.,
105 2009) populations. However, the demand for scanning may lead to novice technicians being
106 utilised and even though these technicians are accredited, they may not have the extended
107 training and experience using the device to attain accurate and reliable results.

108

109 As such, the rationale for this study was that the DXA data being collected by a novice
110 technician was showing high quality results. The question was then asked, how close were the
111 results of a novice technician to those already obtained by an experienced technician utilising
112 the same DXA scanner?

113

114 Therefore, the aim of this study is to ascertain if a novice technician can produce whole body
115 BC and BMD DXA scanning results similar to that of an experienced technician when scanning
116 recreational athletes.

117

118

119 **METHODS**

120 **Study Design**

121 In order to assess the novice and experienced technicians' reliability, precision and sensitivity
122 to change individuals were assigned to a group (experienced or novice). Individuals total body

123 BC and BMD were scanned twice on the same day, minutes apart with repositioning between
124 each scan. Scanning took place in accordance with positioning protocols developed by Alisa
125 Nana as illustrated in Figure 1. (Nana et al., 2012). The study had ethical approval Bond
126 University Human Research Ethics Committee (RO15221, RO1655).

127

128 **Participants**

129 A total of 38 participants were included in this two-part pilot study. Eight participants formed
130 the experienced technicians' group, which was a convenience sample. These eight
131 participants were scanned twice by the same experienced technician to establish their own
132 reliability. The second group (novice technician's group), which consisted of 30 participants,
133 was scanned twice by the same novice technician to determine his reliability. Ethical
134 approval was only granted for individual's to be scanned twice due to concerns over radiation
135 exposure.

136

137 All participants recruited were aged over eighteen, recreational athletes and were from the local
138 geographical area. To be eligible for the study, participants must have been willing to meet
139 scanning stipulations (fasted, bladder voided, removal of metal, abstained from exercise on day
140 of scan and undertake anthropometric assessment). Participants were excluded from the study;
141 if the participant competed in collegiate or professional sport, suspected they were pregnant
142 and or were non-healthy: inclusive of osteoporosis, current fractures, hemiarthroplasty and
143 total joint replacements, rheumatoid or osteoarthritis, current cardiac or pulmonary conditions,
144 diabetes or if they were unable to maintain the required position for the duration of the scans.
145 To reduce the likelihood of artifacts, male participants wore underwear during scanning while
146 female participants wore underwear, sports bra or two-piece bathers. Participants initially were
147 informed of all testing procedures and questions were answered at that time prior to signing

148 the voluntary consent form. No participants who were invited to participate in the study
149 declined to participate. Participants were assessed for height (to the nearest 1.0 cm) using a
150 stadiometer (Harpenden, Holtain Limited, Crymych, UK) and mass (to the nearest 100 grams)
151 using calibrated scales (WM202, Wedderburn, Bilinga, Australia) prior to scanning.

152

153 **Technicians**

154 Both technicians were accredited and trained through the Australia and New Zealand Bone
155 Mineral Society (ANZBMS). The ANZBMS accreditation is the only certification course
156 available which satisfies the requirements of radiation safety legislation in Australia, leading
157 to licensure. Both technicians undertook the same accreditation process. The accreditation
158 course consists of theoretical knowledge and practical skills involved with DXA usage,
159 including bone pathology, device usage, manipulation and the analysis of results. Prior to this
160 study the novice technician's previous experience was approximately 25 DXA scans. The
161 experienced technician was deemed so, as they had completion of more than 100 scans as well
162 as having a five-year history as a DXA technician (Kim et al., 2014).

163

164 **Equipment**

165 All scans were performed using a narrow angle fan beam Lunar Prodigy DXA scanner (GE
166 Healthcare, Madison, WI). Scans were analysed automatically by the GE enCORE 2016
167 software (GE Healthcare). Scans were then analysed by the DXA technician and region of
168 interest lines adjusted accordingly, if needed, relative to the ANZBMS guidelines. The DXA
169 scanner was calibrated daily using a whole body phantom as per manufacturer's guidelines
170 prior to any scans.

171

172 **Statistical Analysis**

173 All data was analysed using IBM statistical package for the social sciences (SPSS, version 24)
174 or via a customised reliability spreadsheet from sportsci.org. To analyse test re-test reliability
175 the recommended Intraclass Correlation Coefficient (3,1) with 95% confidence intervals was
176 performed using SPSS (Ionan et al., 2014; Trevethan, 2016). The ICC results were interpreted
177 as indicators of reliability as follows: ICC of 0.00–0.29, very low reliability; 0.30–0.49, low
178 reliability; 0.50–0.69, moderate reliability; 0.70–0.89, high reliability; and 0.90–1.00, very
179 high reliability (Munro et al., 2005). Additionally, SPSS was used to calculate the standard
180 error of measurement percentage (SEM%) (Equation 1) and smallest real difference percentage
181 (SRD%) (Equation 2) (Lexell et al., 2005). Acceptable precision of results has been previously
182 set by ISCD at 2% for LM and 2% for FM respectively (ISCD, 2015).

$$\text{SEM} = ((\sqrt{\text{mean square error from ANOVA}}/\text{mean}) \times 100, \quad (1)$$

$$\text{SRD}\% = ((1.96 \times \text{SEM} \times \sqrt{2})/\text{mean}) \times 100, \quad (2)$$

183 A customised spreadsheet from Sportscience website (www.sportsci.org) was utilised to
184 calculate and analyse percentage change in mean and the accompanying typical error
185 (coefficient of variation (CV%) percentage) as recommended (Hopkins, 2000; Hopkins et al.,
186 2009).

187

188 **RESULTS**

189 Anthropometrical data (mean + SD) of the participants are presented in Table 1. Independent
190 T-tests for age, height, weight, BMI, whole body FM percentage, whole body LM percentage,
191 whole body BMC percentage and whole body BMD revealed no significant differences
192 between the novice and experienced groups ($p = 0.96, 0.45, 0.21, 0.35, 0.13, 0.06, 0.01, 0.49$
193 respectively) except for BMC.

194

195 All the collated results from the experience and novice technicians' reliability, precision and
196 sensitivity to change are presented in Table 2. Both technicians ICC reliability values were
197 within the high to very high range (Munro et al., 2005)

198 **Experienced technician**

199 Scan 1 produced the following absolute values: FM 23.01%, LM 73.69%, BMC % 3.30%,
200 BMD 1.275 g.cm⁻². Scan 2 produced the following 23.13%, 73.65%, 3.31%, 1.274 g/cm²,
201 difference of 0.12%, 0.04%, 0.01% and 0.001 g/cm² which was evident in the high reliability
202 scores.

203

204 **Novice technician**

205 Scan 1 produced the following absolute values: FM 25.91%, LM 70.21%, BMC % 3.87, BMD
206 1.312 g/cm⁻². Scan 2 produced the following 25.96%, 70.38%, 3.89%, 1.308 g/cm², difference
207 of 0.05%, 0.18%, 0.02% and 0.004 g/cm² which was evident in the high reliability scores.

208

209 **DISCUSSION**

210 The purpose of this study was to ascertain if a novice but accredited DXA technician could
211 produce results similar to that of an experienced DXA technician. DXA reliability has been
212 studied extensively in both the facets of whole body and segmental BC (Bilsborough et al.,
213 2014; Kerr et al., 2016; Nana et al., 2012, 2013) and region specific BMD (Fuller et al., 2016;
214 Lohman et al., 2009). To our knowledge, there is no study to date that has assessed the
215 reliability, precision or sensitivity to change of BC or BMD scanning when completed by a
216 novice technician. Our results indicated that when an accredited, but novice technician uses the
217 Lunar DXA scanner to assess BC and BMD they produce results that are similar, yet slightly
218 inferior to that of an experienced technician.

219

220 The novice and the experienced technician produced very similar fat percentage results. Both
221 technicians achieved very high test-retest reliability (ICC 0.995 and 0.996, and 0.996, 0.10 and
222 0.23 percentage change in mean) and the results are similar to previously published data (ICC
223 0.98 to 0.99, percentage change in mean 0.0 to 0.4) (Bilsborough et al., 2014; Kerr et al., 2016;
224 Nana et al., 2012, 2013). However, the percentage fat parameter produced the worst precision
225 (SEM%) and poorest sensitivity to change (SRD%) statistics compared with the parameters of
226 bone and LM. This is due to the fat parameter producing the largest variance (error rate) of the
227 parameters. This finding of fat tissue producing poorer reliability results is consistent across
228 several BC studies (Bilsborough et al., 2014; Kerr et al., 2016; Nana et al., 2012, 2013), which
229 is then exacerbated when calculating SEM and SRD. Additionally, the novice technicians
230 group had a larger fluctuation in stature of participants with some (n=7) only just fitting within
231 the scanning field, which would have increased the statistical variance. This increase in
232 statistical variance contributed to the experienced technician having better precision (CV%
233 0.33 vs 0.36, SEM% 1.39% vs 2.12) and sensitivity to change (SRD% 3.85% vs 5.86%).
234 However, the precision results (CV%) (0.36% and 0.33% respectively) of the novice and
235 experienced technicians falls well below the range of previously published CV% data of 1.3 to
236 5.9% (Bilsborough et al., 2014; Kerr et al., 2016; Nana et al., 2012, 2013). It should be noted
237 that the sensitivity to change (SEM%) of the experienced technician (1.39%) is well below the
238 ISCD recommend precision (2%) (ISCD, 2015), indicating superior precision, however the
239 SEM% (2.12%) of the novice is just above the recommend precision illustrating that the
240 novice's precision was slightly worse than recommended and may be due to inexperience in
241 positioning and assessing scans, or the larger fluctuation of stature creating higher statistical
242 variance.
243

244 The novice technician had slightly better reliability results when assessing the lean mass
245 parameter (ICC 0.996 vs 0.989, percentage change in mean -0.03 vs -0.10) however the
246 experienced technician demonstrated better precision (CV% 0.47 vs CV% 0.75, SEM% 0.61
247 vs 1.46%) and sensitivity to change (SRD% 1.70 vs 4.05). The reliability of the novice and
248 experienced technician is slightly lower than previously published data when using the ICC
249 statistic (0.996 and 0.989 vs 1.00), however all results are deemed as very high reliability
250 (Munro et al., 2005). When using the percentage change in mean statistic the results are very
251 similar (-0.03 and -0.10 vs range of 0.0 to 0.3.) This fluctuation in reliability results may be
252 due to the type of athlete scanned in the previous studies (professional athletes versus
253 recreational) and the variances in the statistical analysis. When assessing precision the novice
254 and experienced technicians results (CV% 0.75 and 0.47) fall into the lower end of the
255 published data range (0.3 to 1.5%) and the SEM% (0.61 to 1.46%) is well within the ISCD
256 recommendations (2%) (ISCD, 2015), indicating high precision by the technicians in this
257 study.

258
259 When assessing the reliability of BMC% the novice technician produced a slightly higher ICC
260 (0.997 vs 0.994), both of which are deemed as very high (Munro et al., 2005). When comparing
261 the experienced and novice technicians the reliability, when using the ICC statistic is very
262 similar to previously published data (Bilsborough et al., 2014), and at the lower end of the
263 published percentage change in mean (0.02 vs 0.00 – 1.9%). The precision of both technicians
264 is very good with the experienced technician producing slightly better SEM% (0.88% vs 1.1%),
265 which may be due to the smaller sample size of the experienced technician. The precision when
266 expressed as CV% is very low (0.03) in comparison to the large range displayed in previous
267 studies (0.06 – 5.2%) (Bilsborough et al., 2014; Kerr et al., 2016; Nana et al., 2012, 2013),
268 indicating that both the experienced and novice technicians in this study produced very precise

269 results when assessing BMC%. The sensitivity to change of the experienced technician is also
270 lower than the novice technician (2.44% vs 3.10%) indicating better results.

271

272 The reliability of the experienced technician (ICC 0.998) is very high and is clearly more
273 superior to the novice technician's high reliability (0.985) (Munro et al., 2005). Not
274 surprisingly the precision of the experienced technician is also more superior to that of the
275 novice (CV% 0.01 vs 0.14, SEM% 0.70 vs 1.06%). Previously, BMD analysis has been used
276 on site-specific basis i.e. lumbar spine, hip to assess for changes after the occurrence of
277 symptoms, however for this study it was assessed for the entire body as it was being utilised as
278 a screening tool for those at risk of developing bony stress related injuries. As such there has
279 been no reliability data published, however the results of this study (experienced $1.27 + 0.20$
280 $\text{g.cm}^2 \pm 0.11$, inexperienced $1.31 + 0.11 \text{ g.cm}^2$) in terms of grams per centimeter squared are
281 similar to those of previously published data (1.04 ± 0.07 to $1.31 \pm 0.08 \text{ g/cm}^2$) of athletes who
282 are involved in sports that are deemed high risk for stress reactions due to low BMD (Andreoli
283 et al., 2001; Ferry et al., 2011).

284

285 One identified limitation was the use of whole-body BMD measurement as opposed to site-
286 specific BMD measurements because the technology embedded in the BC scan allows for
287 whole body BMD analysis, subsequently reducing the levels of exposure to radiation. For this
288 reason, professional athletes who routinely have BC scan should include a whole-body BMD
289 assessment from the BC scans. The authors recommend that if the whole-body BMD scans
290 results were to show a cause for concern, a segmental site specific BMD can then be
291 undertaken.

292 Furthermore, this study only assessed one experienced technician and one novice technician,
293 using two different sample groups of different sizes, as significant multiple scanning and

294 exposure to radiation was an ethical consideration. Ideally, future research should include
295 multiple technicians scanning large participant cohorts in a cross-sectional design to further
296 validate the findings of this study and minimise the impact of a single technician. To be able
297 to further generalise the findings the sample population should include both recreational and
298 professional athletes.

299

300 In summary, the high to very high reliability results of DXA scanning for both technicians
301 compared with previously published data illustrates that extensive experience whilst valuable
302 is not necessarily a requirement to produce quality results. In a climate where DXA use is
303 becoming a more common place, the results of this study will provide the novice technician
304 with more confidence when completing DXA scanning.

305

306 **ACKNOWLEDGMENTS**

307 This study was granted ethics approval from Bond University Human Research
308 Ethics Committee (RO15221, RO1655). The authors received no funding for this work and
309 have no conflicts of interest. The authors would like to thank the participants for partaking in
310 the study, the graduating Physiotherapy cohort of 2017, and Vini Simas for his time and effort
311 in completing the scans. This study was designed by MC and FS; data were collected by FS
312 and CP; data interpretation and manuscript preparation were undertaken by CP, FS, M.C and
313 J.F. All authors approved the final version of the paper.

314 **REFERENCES**

315 Andreoli, A., Monteleone, M., Van Loan, M., Promenzio, L., Tarantino, U., & De Lorenzo,
316 A. (2001). Effects of different sports on bone density and muscle mass in highly
317 trained athletes. *Medicine & Science in Sports & Exercise*, 33(4), 507 - 511.

- 318 Baniak, N., Grzybowski, S., & Olszynski, W. P. (2014). Dual-Energy X-ray Absorptiometry
319 Scan Autoanalysis vs Manual Analysis. *Journal of Clinical Densitometry*, 17(1), 97-
320 103.
- 321 Bennell, K. L., Malcolm, S. A., Thomas, S. A., Reid, S. J., Brukner, P. D., Ebeling, P. R., &
322 Wark, J. D. (1996). Risk factors for stress fractures in track and field athletes. A
323 twelve-month prospective study. *The American journal of sports medicine*, 24(6),
324 810-818.
- 325 Bilsborough, J. C., Greenway, K., Opar, D., Livingstone, S., Cordy, J., & Coutts, A. J.
326 (2014). The accuracy and precision of DXA for assessing body composition in team
327 sport athletes. *Journal of sports sciences*, 32(19), 1821-1828.
328 doi:10.1080/02640414.2014.926380
- 329 Blake, G. M., & Fogelman, I. (2007). The role of DXA bone density scans in the diagnosis
330 and treatment of osteoporosis. *Postgraduate Medical Journal*, 83(982), 509-517.
331 doi:10.1136/pgmj.2007.057505
- 332 Buckinx, F., Landi, F., Cesari, M., Fielding, R. A., Visser, M., Engelke, K., . . . Kanis, J. A.
333 (2018). Pitfalls in the measurement of muscle mass: a need for a reference standard. *J*
334 *Cachexia Sarcopenia Muscle*, 9(2), 269-278. doi:10.1002/jcsm.12268
- 335 Chen, Y.-T., Tenforde, A. S., & Fredericson, M. (2013). Update on stress fractures in female
336 athletes: epidemiology, treatment, and prevention. *Current Reviews in*
337 *Musculoskeletal Medicine*, 6(2), 173-181. doi:10.1007/s12178-013-9167-x
- 338 Duthie, G. M. (2006). A framework for the physical development of elite rugby union
339 players. *Int J Sports Physiol Perform*, 1(1), 2-13.
- 340 Ferry, B., Duclos, M., Burt, L., Therre, P., Le Gall, F., Jaffré, C., & Courteix, D. (2011).
341 Bone geometry and strength adaptations to physical constraints inherent in different

- 342 sports: comparison between elite female soccer players and swimmers. *Journal of*
343 *Bone and Mineral Metabolism*, 29(3), 342–351.
- 344 Fredericson, M., Ngo, J., & Cobb, K. (2005). Effects of ball sports on future risk of stress
345 fracture in runners. *Clin J Sport Med*, 15(3), 136-141.
- 346 Fuller, J. T., Archer, J., Buckley, J. D., Tsiros, M. D., & Thewlis, D. (2016). The reliability of
347 dual-energy X-ray absorptiometry measurements of bone mineral density in the
348 metatarsals. *Skeletal Radiology*, 45(1), 135-140. doi:10.1007/s00256-015-2227-0
- 349 Georgeson, E. C., Weeks, B. K., McLellan, C. P., & Beck, B. R. (2011). Body Composition
350 Change Over a Professional Rugby League Season and Relationship to Rates and
351 Types of Injury. *Medicine & Science in Sports & Exercise*, 43(5 (Suppl 1)), 108.
- 352 Hagmar, M., Berglund, B., Brismar, K., & Hirschberg, A. L. (2013). Body composition and
353 endocrine profile of male Olympic athletes striving for leanness. *Clin J Sport Med*,
354 23(3), 197-201. doi:10.1097/JSM.0b013e31827a8809
- 355 Hangartner, T. N., Warner, S., Braillon, P., Jankowski, L., & Shepherd, J. (2013). The
356 Official Positions of the International Society for Clinical Densitometry: acquisition
357 of dual-energy X-ray absorptiometry body composition and considerations regarding
358 analysis and repeatability of measures. *Journal of Clinical Densitometry*, 16(4), 520-
359 536. doi:10.1016/j.jocd.2013.08.007
- 360 Hopkins, W. G. (2000). Measures of Reliability in Sports Medicine and Science. *Sports*
361 *Medicine*, 30(1), 1-15. doi:10.2165/00007256-200030010-00001
- 362 Hopkins, W. G., Marshall, S. W., Batterham, A. M., & Hanin, J. (2009). Progressive statistics
363 for studies in sports medicine and exercise science. *Medicine & Science in Sports &*
364 *Exercise*, 41(1), 3-13. doi:10.1249/MSS.0b013e31818cb278
- 365 International Society Clinical Densitometry (ISCD). (2015). *ISCD 2015 Adult Official*
366 *Positions*. Middletown, CT, USA: ISCD.

- 367 Ionan, A. C., Polley, M.-Y. C., McShane, L. M., & Dobbin, K. K. (2014). Comparison of
368 confidence interval methods for an intra-class correlation coefficient (ICC). *BMC*
369 *Medical Research Methodology*, *14*, 121. doi:10.1186/1471-2288-14-121
- 370 Johnell, O., & Kanis, J. A. (2006). An estimate of the worldwide prevalence and disability
371 associated with osteoporotic fractures. *Osteoporos Int*, *17*(12), 1726-1733.
372 doi:10.1007/s00198-006-0172-4
- 373 Kelsey, J. L., Bachrach, L. K., Procter-Gray, E., Nieves, J., Greendale, G. A., Sowers, M., . . .
374 Cobb, K. L. (2007). Risk factors for stress fracture among young female cross-
375 country runners. *Medicine & Science in Sports & Exercise*, *39*(9), 1457-1463.
376 doi:10.1249/mss.0b013e318074e54b
- 377 Kerr, A., Slater, G. J., Byrne, N., & Nana, A. (2016). Reliability of 2 Different Positioning
378 Protocols for Dual-Energy X-ray Absorptiometry Measurement of Body Composition
379 in Healthy Adults. *Journal of Clinical Densitometry*, *19*(3), 282-289.
380 doi:10.1016/j.jocd.2015.08.002
- 381 Kiebzak, G. M., Leamy, L. J., Pierson, L. M., Nord, R. H., & Zhang, Z. Y. (2000).
382 Measurement precision of body composition variables using the lunar DPX-L
383 densitometer. *Journal of Clinical Densitometry*, *3*(1), 35-41.
- 384 Kim, H. S., & Yang, S. O. (2014). Quality control of DXA system and precision test of radio-
385 technologists. *J Bone Metab*, *21*, 2-7.
- 386 Lewiecki, E. M. (2005). Clinical applications of bone density testing for osteoporosis.
387 *Minerva medica*, *96*(5), 317-330.
- 388 Lexell, J. E., & Downham, D. Y. (2005). How to assess the reliability of measurements in
389 rehabilitation. *Am J Phys Med Rehabil*, *84*(9), 719-723.
- 390 Lohman, M., Tallroth, K., Kettunen, J. A., & Marttinen, M. T. (2009). Reproducibility of
391 dual-energy x-ray absorptiometry total and regional body composition measurements

- 392 using different scanning positions and definitions of regions. *Metabolism: clinical and*
393 *experimental*, 58(11), 1663-1668. doi:10.1016/j.metabol.2009.05.023
- 394 Munro, B., & Visintainer, M. (2005). *Statistical methods for health care research* (5th ed.
395 Vol. 1): Lippincott, Philadelphia.
- 396 Nana, A., Slater, G. J., Hopkins, W. G., & Burke, L. M. (2012). Effects of daily activities on
397 dual-energy X-ray absorptiometry measurements of body composition in active
398 people. *Medicine & Science in Sports & Exercise*, 44(1), 180-189.
399 doi:10.1249/MSS.0b013e318228b60e
- 400 Nana, A., Slater, G. J., Hopkins, W. G., & Burke, L. M. (2013). Effects of exercise sessions
401 on DXA measurements of body composition in active people. *Medicine & Science in*
402 *Sports & Exercise*, 45(1), 178-185. doi:10.1249/MSS.0b013e31826c9cfd
- 403 Newton, R., Taaffe, D., Spry, N., Gardiner, R., Levin, G., Wall, B., . . . Galvao, D. (2009). A
404 phase III clinical trial of exercise modalities on treatment side-effects in men
405 receiving therapy for prostate cancer. *BMC Cancer*, 9, 210.
- 406 Schnackenburg, K. E., Macdonald, H. M., Ferber, R., Wiley, J. P., & Boyd, S. K. (2011).
407 Bone quality and muscle strength in female athletes with lower limb stress fractures.
408 *Medicine & Science in Sports & Exercise*, 43(11), 2110-2119.
409 doi:10.1249/MSS.0b013e31821f8634
- 410 Scofield, K. L., & Hecht, S. (2012). Bone health in endurance athletes: runners, cyclists, and
411 swimmers. *Curr Sports Med Rep*, 11(6), 328-334.
412 doi:10.1249/JSR.0b013e3182779193
- 413 Standorth, D., Stults-Kolemainen, M., & Stanforth, P. (2016). Bone Mineral Content and
414 Density Among Female NCAA Division I Athletes Across the Competitive Season
415 and Over a Multi-Year Time Frame. . *Journal of Strength & Conditioning Research*,
416 30(10), 2828-2838.

- 417 Stewart, A. D. (2001). Assessing body composition in athletes. *Nutrition*, 17(7-8), 694-695.
- 418 Tenforde, A. S., Sainani, K. L., Carter Sayres, L., Milgrom, C., & Fredericson, M. (2015).
 419 Participation in ball sports may represent a prehabilitation strategy to prevent future
 420 stress fractures and promote bone health in young athletes. *Physical Therapy*, 7(2), 222-225.
 421 doi:10.1016/j.pmrj.2014.09.017
- 422 Trevethan, R. (2016). Intraclass correlation coefficients: clearing the air, extending some
 423 cautions, and making some requests. *Health Services and Outcomes Research
 424 Methodology*, 1-17. doi:10.1007/s10742-016-0156-6
- 425 Trexler, E., Smith-Ryan, A., Mann, J., Ivey, P., Hirsch, K., & Mock, M. (2018). Longitudinal
 426 body composition changes in NCAA division I college football players. *Journal of
 427 Strength & Conditioning Research*, 31(1), 1-8.
- 428 Watts, J.J.; Abimanyi-Ochom, J.; Sander, K.M. (2013). *Osteoporosis costing all Australians.
 429 A new burden of disease analysis – 2012 to 2022: Osteoporosis Australia.*

430

431

432 **TABLES**433 **Table 1.** Demographical Data

| | Age (years) | Height (cm) | Mass (kgs) | BMI (range) | FM % | LM % | BMC% | BMD (g/cm ²) |
|-------------|----------------|-----------------|----------------|-----------------|-----------------|-----------------|----------------|-----------------------------|
| Experienced | 29.2 ± 11.5 | 175.8 ± 2.6 | 78.75 ± 6.9 | 23.0 to 29.2 | 23.07 ± 4.49 | 73.62 ± 4.30 | 3.31 ± 0.51 | 1.27 ± 0.19 |
| Novice | 29.6 ± 10.0 | 171.7 ± 10.7 | 70.6 ± 12.4 | 19.4 to 31.7 | 26.03 ± 7.29 | 70.07 ± 7.03 | 3.89 ± 0.45 | 1.31 ± 0.11 |

434 cm – centimetres, kgs – kilograms, FM % - Fat mass percentage, LM % - Lean mass percentage, BMC% - bone
 435 mineral content percentage, BMD – bone mineral density, g/cm² – grams per centimetre squared

436

437 **Table 2.** Reliability, precision and sensitivity to change results for experienced and
 438 novice technicians.

| | | ICC | CI | % Δ in Mean | CV% | SEM% | SRD% |
|-------------|-----------------------|-------|---------------|--------------------|------|-------|-------|
| Experienced | Fat % | 0.995 | 0.976 – 0.999 | 0.10 | 0.33 | 1.39% | 3.85% |
| | Lean % | 0.989 | 0.949 – 0.998 | -0.10 | 0.47 | 0.61% | 1.70% |
| | BMC (g) | 0.994 | 0.973 – 0.999 | 0.02 | 0.03 | 0.88% | 2.44% |
| | BMD g/cm ² | 0.998 | 0.991 – 1.000 | -0.01 | 0.01 | 0.70% | 1.90% |
| Novice | Fat % | 0.996 | 0.990 – 0.998 | 0.23 | 0.36 | 2.12% | 5.86% |
| | Lean % | 0.996 | 0.991 – 0.998 | -0.03 | 0.75 | 1.46% | 4.05% |
| | BMC (g) | 0.997 | 0.993 – 0.999 | 0.02 | 0.03 | 1.10% | 3.10% |
| | BMD g/cm ² | 0.985 | 0.970 – 0.993 | -0.04 | 0.14 | 1.06% | 2.73% |

439 g/cm² – grams per centimetre squared, % Δ in Mean – percentage change in mean, CV- confidence variance (typical error),
 440 ICC – intraclass correlation coefficient, CI – confidence interval, SEM% - percentage standard error of measurement, SRD%
 441 - percentage smallest real difference

442

443

444

445 **FIGURES**

446

447 **Figure 1.** Positioning Protocol.

448

449